



कर्णाली प्रदेश सरकार
सामाजिक विकास मन्त्रालय

फोन नं. ०८३ ५२३२००
फ्याक्स ०८३ ५२०२६६

प्रदेश अस्पताल वीरेन्द्रनगर, सुर्खेत

एम्बुलेन्स खरिद सम्बन्धि सूचना दोस्रो प्रकासित मिति २०७९/०९/२६

यस अस्पताललाई देहाय बमोजिमको “ख” वर्गको एम्बुलेन्स सार्वजनिक खरिद ऐन २०६३ को दफा ८ को उपदफा (१) (क) (८) तथा सार्वजनिक खरिद नियमावली २०६४ को नियम ३१ ख बमोजिम त्यस्तो एम्बुलेन्स उत्पादन कम्पनी वा सोको आधिकारिक विक्रेताहरु बिच प्रतिस्पर्धा गराउने (क्याटलग सपिड) विधिबाट खरिद गर्नु पर्ने भएकोले उक्त नियमावलीको नियम ३१ (ख) को उपनियम २ अनुसार देहाय बमोजिमको एम्बुलेन्स बिक्री गर्न इच्छुक उत्पादक वा आधिकारिक विक्रेताले देहाय बमोजिमका कागजात सहित सूचना प्रकासित भएको मितिले १० औं दिनको ४:३० बजे भित्र स्पेसिफिकेसन, गुणस्तर, मूल्य र सुविधा सहितको विवरण (क्याटलग वा ब्रोसर) संलग्न राखी यस अस्पतालमा सिलबन्दि प्रस्ताव दर्ता गर्नु हुन सम्बन्धित उत्पादक वा आधिकारिक विक्रेतालाई जानकारीको लागि यो सूचना प्रकासित गरिएको छ । विस्तृत जानकारीको लागि यस अस्पतालमा सम्पर्क गर्न सकिनेछ ।

प्रस्तावदाताले पेश गर्नु पर्ने आवश्यक कागजातहरु:

- नवीकरण गरेको फर्म वाकम्पनी दर्ता प्रमाणपत्रको प्रतिलिपी ।
- मूल्य अभिवृद्धि कर (VAT) दर्ता प्रमाण पत्रको प्रतिलिपी
- आ.व. ०७७/०७८ को करचुकता प्रमाणपत्रको प्रतिलिपी ।
- खरिद कारवाहिमा नपरेको स्व:घोषणापत्र ।
- एम्बुलेन्स तथा उपकरणको Technical Specification प्रमाणित गर्ने रंगिन ब्रोसर/Catalogue or Technical data sheet.
- मूल्य खुल्ने प्रमाणित कागजात (पत्रिकामा प्रकासित भएको वा कम्पनीको लेटरप्याडमा आधिकारिक व्यक्तिले सहिछाप गरी प्रमाणित गरेको कागजात)।
- Warranty or Guarantee सम्बन्धि कागजात ।
- अस्पताललाई सम्बोधन गरिएको Manufacturer's Authorization Letter as per Specification.
- अस्पतालको web site: www.provincehospital.karnali.gov.np मा राखिएको Technical Specification डाउनलोड गरि भरेर सहिछाप सहित पेश गर्नु पर्नेछ ।

“ख” वर्गको एम्बुलेन्सको विवरण

Description	General Specification
Ambulance 4WD category "B" No.1.	4 WD, Right Hand Drive model, Right Hand Drive model suitable for driving in Nepal, Diesel Engine, Displacement Capacity not less than 2550 CC, Max. Power output shall not be less than 110 KW & Max. Torque Shall not be less than 350 Nm @ governed RPM, Dimension: Overall Length Min. 5000mm, Overall width Min.1500mm, Minimum ground clearance Should be 200 mm. Transmission; Manual Shift at least 5 Forward and One Reverse or automatic transmission. Brakes; Front ventilated Disc type, rear disc / drum type, hydraulic operated on all wheels, and parking brake, dual circuit preferred.. Fuel Tank; Fuel tank capacity should be min, 70ltrs. All interior structure reinforced with treatment anti-corrosion, Air conditioning System on the rear compartment. Should have cabinets and drawers for storing all medical parts with anti-accidental-opening locks & All Medical Equipment for Ambulance category "B". For further details Please download the Technical Specification of Ambulance category "B" from the hospital web site; www.provincehospital.karnali.gov.np .

अस्पताल निर्देशक

TECHNICAL SPECIFICATIONS OF AMBULANCE FOR CATEGORY "B" TYPE

S.N.	PARTICULARS	REQUIREMENT	BIDDERS OFFER
1	General	Pickup Ambulance, Box design, 4WD, Right Hand Drive model suitable for driving in Nepal. Weight of the Vehicle shall not be less than 2800 kgs. The Vehicle shall be comfortable space for medical equipment (As per Accessories list & Specifications), handling staffs, Driver, Co-Driver, Patient & Care Taker person. Vehicle shall be latest model.	Make: Model: Gross Weight: Mfg. Year Country of Origin:
2	Engine	Diesel Engine, Displacement Capacity not less than 2550 CC, Max. Power output shall not be less than 110 KW & Max. Torque Shall not be less than 350 Nm @ governed RPM. Engine fully equipped with accessories for efficient full load operation of the vehicle at operating conditions.	Make: Model: Engine: Displacement cc.
3	Emission Standard	To emission standard of the offered vehicle must be in compliance with of superior then the applicable emission standard prescribed by Nepal vehicle mass commission standard Which is to be certified by department of transport management Nepal	
4	Gauges	To include water temperature, fuel, oil pressure, transmission gauge etc. as necessary for efficient operation and maintenance.	
5	Dimension	Overall Length Min. 5000mm Overall width Min.1500mm Minimum ground clearance Should be 200 mm	
6	Transmission	Manual Shift at least 5 Forward and One Reverse or automatic transmission.	
7	Wheel Base	Minimum 3000mm	
8	Tires	Standard size	
9	Steering	Power Steering	
10	Brakes	Front ventilated Disc type, rear disc / drum type, hydraulic operated on all wheels, and parking brake, dual circuit preferred.	
11	Front cab	All steel cab with insulated roof laminated glass Windscreen. Wiper, glass windows on doors. proper ventilated system. Comfortable seats with adjustable seat for driver, two sun visors, two rear view mirror, Seat belt for driver and co-driver.	
12	Patient cab	At Least 1 back bi-fold door, all metallic cab with insulated roof, laminated glass windscreen, front washer and wipers, glass windows on doors, thorough ventilation, comfortable and adjustable seats for co- passenger. Sun visor, rear view mirror and two side view mirrors. -All interior structure reinforced with treatment anti-corrosion -Air conditioning System on the rear compartment. -Should have cabinets and drawers for storing all medical parts with anti-accidental-opening locks.	
13	Fuel Tank	Fuel tank capacity should be min, 70 ltrs	
14	Essential accessories	<ol style="list-style-type: none"> 1. Head light, tail and rear light, reversing light, and turn signals. 2. Cabin Lights 3. Speedometer / Odometer. 4. One Spare wheel with tyre and tube. 5. Suitable jack and handle. 6. HVAC (Heating & Cooling System) 7. CD / MP 3 Player/Radio AM/FM 	

		<p>8. Touch Screen with reverse Camera</p> <p>9. Equipped with 5 seats</p> <p>10. Wheel chair and foldable stretcher</p> <p>11. Siren</p> <p>12. Flashing Light on top (As Per Nepal Standard)</p> <p>13. Oxygen Cylinder holder</p> <p>14. GPS System</p> <p>15. Anti-skid floor</p> <p>16. Radio communication system</p> <p>17. Fire extinguisher</p> <p>18. Dual AirBag</p> <p>19. ABS + EBD</p> <p>Equipped with all accessories offered in the standard model of category "B " type ambulance but not limited</p>	
15	Instructions	All signs and instructions in the vehicle shall be English or Nepali.	
	Proven Performance	The vehicle offered shall be under standard production by the manufacturer for at least five years.	
16	Manuals	One copy of the operators and owners' instructions in English shall be supplied with each vehicle	
17	Tools	A set of tools as required for general maintenance shall be supplied.	
18	Color	2/3 Outer body portion yellow & remaining left/right green & red in Battenberg or chess board design	
19	Warranty	Manufacturer shall provide a minimum of one-year WARRANTY after acceptance.	
20	Delivery	The ambulance [4x4] one unit shall be delivered to: Karnali Province Hospital, Birendranagar Surkhet.	

LIST & SPECIFICATION OF BIOMEDICAL EQUIPMENT.

S/N	PARTICULARS	SPECIFICATIONS	BIDDERS OFFER
1	Portable Suction Unit with Battery Back-up	<ol style="list-style-type: none"> Equipment shall be with 12V DC Maximum negative pressure from approx. -200 to 600 mbar in steps of 100 or less with suitable setting marks. Sufficient capacity 500ml secretion bottles with efficient overflow protected with adjustable negative pressure (Min. 5 Nos. Polycarbonate & autoclave able). Ambulance Wall countable. Power Pack rechargeable Nico Battery 12 V with battery chargers connecting cable for connection to 230v AC+ 10%, 50 Hz and with provision for recharging from the vehicle with vehicle circuit connecting cable. Battery charged life shall be of min. 30minutes. Suction capacity of minimum 5-8 litre per minutes. ISO 13485 or CE Certified 1 year's warranty form the date of acceptance. 	
2	Patient Monitor-5 para	<ol style="list-style-type: none"> 12-inch LED/LCD color screen with Touch Screen. at least 7 wave form simultaneously display. 3 lead ECG, NIBP, SpO2, 2-Temp., Resp, HR. ISO 13485 /CE / USFDA Certified 1 year's warranty form the date of acceptance. 	
3	Manual BP Set	<ol style="list-style-type: none"> Manual BP set with accessories Measuring range:40 – 200 mm Hg. ISO 13485/ISO 9001 Certified 	
4	ECG 3 Channel	<ol style="list-style-type: none"> At least 5-inch LCD display having touch screen operability. ISO 13485 / CE Certified. 1 year's warranty form the date of acceptance. 	
5	Electric Nebulizer Machine	<ol style="list-style-type: none"> With necessary accessories- nebulization mask, tubing and cable cord for nebulizer Compressor type ISO 13485/ISO 9001 Certified 	
6	Handheld Pulse Oximeter	<ol style="list-style-type: none"> Should consists of color TFT display with multi directional views with brightness control Handheld pulse oximeter with stand type charger with adopter SpO2, HR or PI ISO 13485/CE Certified 	
7	CPR Board	<ol style="list-style-type: none"> Red/Blue Color ISO 13485/ISO 9001 Certified 	
8	Intubation Kit	<ol style="list-style-type: none"> Laryngoscope Handle: 2 Nos. Laryngoscope Blade: 3 Nos. complete set- i.e. Small, Medium and Large Guedel airway set (0,1,2,3,4): 1 No Endotracheal Tube set (6,7,8,9): complete set– plain i.e. non cuffed and cuffed Adhesive Tape Laryngeal Mask airways LMA: 1 No. Proseal 3,4,5 Nos. ISO 13485/ISO 9001 Certified 	

9	Fully automated external defibrillator (AED)	<ol style="list-style-type: none"> Unit should be lightweight compact and portable (not exceeding 6kg.) Unit should have facility for Automatic External Defibrillation and manual defibrillation. Controlled biphasic wave form and up to 200 shocks applicable with a new battery. Should have visual and audible alarm indicators. Should have battery backup at least for 30 discharges of 200J or more and regular monitoring of battery levels. Should have extra disposable self-sticking electrodes/pads- 10set each USFDA. /European CE approved model should be offered. 1 year's warranty form the date of acceptance. 	
10	Laryngoscope set with blades	<ol style="list-style-type: none"> Complete one set with Standard equipment in stainless Steel Laryngoscope with blades Standard equipment in metal with at least 3 standard size curved blades and one extra-large blade (Adult &Child). Handle should have comfortable grip. Good quality light source (Fibre optic/conventional). ISO 13485/ISO 9001 Certified 	
11	Cervical Collar	<ol style="list-style-type: none"> Should be adjustable to 4 different sizes. Should be pre-molded chin support, locking dips and rear ventilation panel, enlarged trachea opening. Should be high-density polyethylene and foam padding with one-piece design enables efficient storage where space is limited. Should be X-ray lucent and easy to clean and disinfect. ISO 13485/ISO 9001 Certified 	
12	Foldable Carrying Chair (Wheel Chair cum Stair Chair)	<ol style="list-style-type: none"> Net weight: less than 10 Kgs Pull through, telescoping long handles built in to lift patients & carry them through narrow passages. Should be as per CE/FDA standards. 	
13	Resuscitation Set (Adult, pediatric and neonatal) – Each 1 set for Adult, pediatric and neonatal	<ol style="list-style-type: none"> Easy Grip manual resuscitator with transparent face mask. Adult models (1500 to 2000ml bag capacity) or better Child models (500 o 250ml bag capacity) or better. Standard 15-22 mm Swivel connector allows connections to all common masks Endotracheal Tubes. Provision to give supplemented oxygen-by oxygen reservoir providing 100%oxygen. Non-rebreathing valve enabling the patient to inspire oxygen from the reservoir bag. ISO 13485/ISO 9001 Certified 	
14	Cylinder, Oxygen, Approx 20L	<ol style="list-style-type: none"> A pressure regulator/flow meter capable of reducing the pressure to appropriate level to run either a ventilator or provide oxygen therapy and oxygen opener. ISO 13485/ISO 9001 Certified 	
15	Splints set of 6 adult sizes with carrying case 1) Hand &Wrist 2) Half arm 3) Full arm 4) Foot and ankle	<ol style="list-style-type: none"> X-ray through the splints Inflatory tubes extension with dosing damp makes dosing easy and quick after inflation. Fixing of splint is by zipper or belt. Distal end left open to expose toes. 5) Should be washable and reusable. 	

	5) Half leg 6) Full leg		
16	Delivery Complete Set	<ol style="list-style-type: none"> 1. Bowl (Small Size 200-250ml) 2 Pcs 2. Bowl (Big Size 1000ml) 1 Pcs 3. Sponge Holding Forceps 10"- 1 Pcs 4. Artery Forceps Straight (6", 8" & 10") 3 Pcs 5. Suture Cutting Scissor 1 Pc 6. Rubber sheet- 1 pcs 7. Abdomen towel-1 pcs 8. Perineal towel 1 Pcs 9. Baby wrapper-2 Pcs 10. Double wrapper – 1 Pcs 11. Abdomen massage towel- 1 Pcs 12. Cord Cutting Scissor (Umbilicus-bliunt)- 1 Pcs 13. ISO 13485/ISO 9001 Certified 	
17	Glucometer	<ol style="list-style-type: none"> 1. The glucometer with test strips and standard 2. ISO 13485/ISO 9001 Certified 	
18	Stethoscope	<ol style="list-style-type: none"> 1. Stethoscope (3M) with standard adult size, chromium plated metal binaural, V rubber tube in one piece. Rotating piper fitting for both functions. (Reputed company make/approved) 2. ISO 13485/ISO 9001 Certified 	
19	Dressing sets	<ol style="list-style-type: none"> 1. Mayo Scissors 14.5cm Straight TC 2. Dissecting Scissors 14.5cm Sharp/Blunt 3. Lister Bandage Scissors 15cm 4. Spencer Stitch Scissors 11cm 5. Foerster Sponge Forceps 25.5cm 6. Dressing Forceps 16cm 7. Dressing Forceps 1:2 16cm 8. Kelly Forceps 14cm Straight 9. Adson Forceps 12cm 10. Lotion Bowl 04 Oz 11. Kidney Dish 08" 12. ISO 13485/ISO 9001 Certified 	
20	Tounge depressure		
21	Torch Light		
22	Hemostatics set		
23	Cauterization sets		

Note:

1. **The major items are indicated by bold and highlighted letters /alphabets/figure in the technical specifications. Non-compliance of these major items will result in non-responsiveness of the bid offered. Bidder's must submit same/ similar specific experience.**
2. **Verification of Technical Data: All technical data supplied by bidder should be verified with manufacturer's authorized website. Bidder should be mention the manufacturer's authorized Web Address (Mandatory).**

Bidder's Signature.....

Date.....

Mail Address/ Contact No.

Seal: